



Acumen Laboratory and Testing Limited

Workshop 04, 7/F, The Whitney, No. 183 Wai Yip Street, Kwun Tong, Kowloon

Tel: (852) 2333 6823 Fax: (852) 2333 1316

Page 1 of 1

Appendix - Quality Control Summary Table

Project Name: Hung Shui Kiu/Ha Tsuen New Development Area Stage 1 Works

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
01/09/2025	R252012	0.22	0.07	3.22	3.12	3.15	10	94.1	Pass
03/09/2025	R252045	0.22	0.08	4.58	4.74	-3.43	10	92.2	Pass
06/09/2025	R252083	0.22	0.09	4.14	3.94	4.95	10	94.0	Pass
10/09/2025	R252107	0.22	0.11	4.66	4.81	-3.17	10	93.0	Pass
12/09/2025	R252112	0.22	0.10	4.79	4.65	2.97	10	93.7	Pass
15/09/2025	R252121	0.22	0.08	4.05	4.20	-3.64	10	94.3	Pass
18/09/2025	R252141	0.22	0.08	4.56	4.36	4.48	10	94.2	Pass
20/09/2025	R252147	0.22	0.09	4.21	4.39	-4.19	10	92.8	Pass
22/09/2025	R252171	0.22	0.10	4.76	4.64	2.55	10	93.5	Pass
26/09/2025	R252193	0.22	0.07	4.57	4.70	-2.80	10	92.7	Pass
29/09/2025	R252210	0.22	0.09	3.36	3.25	3.33	10	93.7	Pass